

(1) The approximate volume of plasma from which a sample unit of Platelets is prepared.

(2) Instructions to begin administration as soon as possible, but not more than 4 hours after entering the container.

(m) For Plasma, the instruction circular shall contain:

(1) A warning against further processing of the frozen product if there is evidence of breakage or thawing.

(2) Instructions to thaw the frozen product at a temperature between 30 and 37 °C.

(3) When applicable, instructions to begin administration of the product within 6 hours after thawing.

(4) Instructions to administer to ABO-group-compatible recipients.

(5) A statement that this product has the same hepatitis risk as Whole Blood; other plasma volume expanders without this risk are available for treating hypovolemia.

(n) For Cryoprecipitated AHF, the instruction circular shall contain:

(1) A statement that the average potency is 80 or more International Units of antihemophilic factor.

(2) The statement: "Usually contains at least 150 milligrams of fibrinogen"; or, alternatively, the average fibrinogen level determined by assay of representative units.

(3) A warning against further processing of the product if there is evidence of breakage or thawing.

(4) Instructions to thaw the product for no more than 15 minutes at a temperature of 37 °C.

(5) Instructions to store at room temperature after thawing and to begin administration as soon as possible but no more than 4 hours after entering the container or after pooling and within 6 hours after thawing.

(6) A statement that 0.9 percent Sodium Chloride Injection U.S.P. is the preferred diluent.

(7) Adequate instructions for pooling to ensure complete removal of all concentrated material from each container.

(8) The statement: "Good patient management requires monitoring treatment responses to Cryoprecipitated AHF transfusions with periodic plasma factor VIII or

fibrinogen assays in hemophilia A and hypofibrinogenemic recipients, respectively."

[50 FR 35470, Aug. 30, 1985, as amended at 53 FR 116, Jan. 5, 1988]

EFFECTIVE DATE NOTE: The information collection requirements contained in §606.122 will not become effective until OMB approval has been obtained. FDA will publish a notice of OMB approval in the FEDERAL REGISTER.

Subpart H—Laboratory Controls

§ 606.140 Laboratory controls.

Laboratory control procedures shall include:

(a) The establishment of scientifically sound and appropriate specifications, standards and test procedures to assure that blood and blood components are safe, pure, potent and effective.

(b) Adequate provisions for monitoring the reliability, accuracy, precision and performance of laboratory test procedures and instruments.

(c) Adequate identification and handling of all test samples so that they are accurately related to the specific unit of product being tested, or to its donor, or to the specific recipient, where applicable.

§ 606.151 Compatibility testing.

Standard operating procedures for compatibility testing shall include the following:

(a) A method of collecting and identifying the blood samples of recipients to ensure positive identification.

(b) The use of fresh recipient serum samples less than 48 hours old for all pretransfusion testing.

(c) The testing of the donor's cells with the recipient's serum (major crossmatch) by a method that will demonstrate agglutinating, coating and hemolytic antibodies, which shall include the antiglobulin method.

(d) A provision that, if the unit of donor's blood has not been screened by a method that will demonstrate agglutinating, coating and hemolytic antibodies, the recipient's cells shall be tested with the donor's serum (minor crossmatch) by a method that will so demonstrate.